



## CE DECLARATION OF CONFORMITY

Micromeritics Instrument Corporation, Inc.  
4356 Communications Dr.  
Norcross, Georgia 30093-2901 U.S.A.  
ISO 9001:2000 Certified

### VERIFICATION

**Model Number:** Gemini VII 2390  
**Type of Equipment:** Physisorption Analyzer  
**Applicable Directive:** 2004/108/EEC European Union (EU) Emissions and Immunity  
73/23/EEC Low Voltage Directive European Union (EU) Safety

**Standards of Conformity:** EN 61326:1997+A1:1998+A2:2001  
EN 61000-4-2:95 Electronic Discharge  
EN 61000-4-3:96 RF Electromagnetic Field  
EN 61000-4-4:95 Electrical Fast Transient/Burst  
EN 61000-4-5:95 Surge  
EN 61000-4-6:96 RF Common Mode  
EN 61000-4-11:1994 Voltage Dips, Interruptions, and Variations  
EN 61326 (CE) Radiated Emissions  
EN 61326 (CE) Conducted Emissions  
EN 61000-3-2:1995+A14:1998 Harmonic Current Emissions  
EN 61000-3-3:1995 Voltage Fluctuations  
  
EN 61010-1 Part 1: General Safety Requirements

**European Contact Address:** Micromeritics Germany GmbH, Avantis Science Park, Rutherford 108, D-52072 Aachen  
Tel: +49 (0) 241 189 446 0, Fax: +49 (0) 241 189 446 11  
email:micromeritics.de@micromeritics.com

Micromeritics Instrument Corporation hereby declares that the instrument(s) specified above conform(s) to the Directive(s) and Standard(s), when installed and operated in accordance with the specifications set forth by Micromeritics. The original of this document is kept at Micromeritics with copies of the relevant test data and certificates, which constitute the required technical file for self declaration.

Uth Khan  
Signature

3-28-2011  
Date

Stefan Koch  
Print name

Director of Engineering  
Position